



# Calibration Certificate Optical

**On site Calibration**  
Certificate nr: **DOA75012402-BY73** (rev.1)  
Driftcon® Service License 2

## Applicant

Institute / Client	Cyclertest
Department	training and support
Address	Rotcherweg 61
Zip / City	6374XE Ladngaarf
Country	Netherlands

## On site location

Contact	Support
Address	Rotscherweg 61
Room	demo room 1 floor
Zip / City	6374 XW Landgraaf
Country	Netherlands
Phone	+31455338733
E-mail	

## Tested instrument (DUT)

Manufacturer	Bio-Rad Laboratories
Instrument	CFX96 Opus Deepwell
Instrument serial number	CT011239
Block type	96-Deep Well
Block position	Center
Block serial number	RD031799
Optical unit serial number	790BR01182
Internal reference	cycler_07
Software version	3.1.1517.0823

## Calibration method

The optical calibration is performed conform ISO 20836:2021 guidelines.

15 LED's (light emission diode) and one temperature sensor are position on top of the calibration unit. The LED emit light at prescribed points in the sample block of the real-time thermal cycler. The emission intensity of light is independent of the measured temperature (Cq determination).

In addition, 19 temperature sensors measure the temperature whereas 15 of them are at the same LED prescribed points in the sample block of the qPCR instrument and are monitored through an ITS-90 traceable measurement system. The sensors are positioned at the bottom of the measurement well. Measurements are made on all temperature sensors simultaneously at 4 Hz. This physical patented Optical calibration method (CYCLERtest BV) is at the moment not fully traceability to SI however offers a substantially higher level of traceability than other methods. (conform ISO 20836:2021)

Traceability SI unit: W/sr (radiant intensity)



### **Thermal cycler protocol**

#### **Pre-heat**

- Step 1 : 30 °C for 60 seconds
- Step 2 : 95 °C for 60 seconds
- Step 3 : 30 °C for 60 seconds

#### **Temperature Calibration**

- Step 1 : 30 °C for 60 seconds
- Step 2 : 95 °C for 60 seconds
- Step 3 : 30 °C for 60 seconds
- Step 4 : 90 °C for 60 seconds
- Step 5 : 50 °C for 60 seconds
- Step 6 : 70 °C for 60 seconds
- Step 7 : 60 °C for 60 seconds
- Step 8 : 30 °C for 60 seconds

#### **Optical calibration – amplification curve**

- Step 1 : 80 °C for 1 second
  - Step 2 : 60 °C for 30 seconds including detection
- Repeat 24 times

### **Optical calibration system protocol**

#### **Pre-heat**

- No light emitted

#### **Temperature Calibration**

- No light emitted

#### **Optical calibration – amplification curve**

- Light intensity temperature independent
- Light intensity gradually increased from 20% to 100% cycle depending



**Non-conformities in calibration method or environmental conditions**

There were no non-conformities reported.

**Environmental conditions (average)**

Ambient temperature	N/A
Relative humidity	N/A

**Notes**

Baseline Threshold Settings: Auto calculated (34.02 from DUT)

**Results**

The calibration results are given on following pages of this certificate.

**Calibration system uncertainty**

The reported uncertainty ( $U$ ) is based on a standard uncertainty multiplied by a coverage factor of  $k=2$ , which provides a confidence level of approximately 95%. The standard uncertainty has been determined in accordance with EA-4/02M.

**Calibration date**

15-Jun-2026

**Due date**

June 2028

**Report date**


19-Jun-2026 08:54

Approved signatory

Mr. Tom Hendrikx  
CEO

Reproduction of the complete certificate is allowed. Parts of the certificate may only be reproduced with written approval of the calibration laboratory. This certificate is issued provided that CYCLERtest BV assumes no liability. The used method and technology, hardware and software are covered by CYCLERtest global patent #2581728.

Temperature sensor/LED positions



Sensor 1 in Well (A1)  
Sensor 2 in Well (D1)  
Sensor 3 in Well (H1)  
Sensor 4 in Well (A4)  
Sensor 5 in Well (E4)  
Sensor 6 in Well (H4)  
Sensor 7 in Well (A7)

Sensor 8 in Well (D7)  
Sensor 9 in Well (H7)  
Sensor 10 in Well (A10)  
Sensor 11 in Well (E10)  
Sensor 12 in Well (H10)  
Sensor 13 in Well (A12)  
Sensor 14 in Well (D12)  
Sensor 15 in Well (H12)

**Heated lid**


Heated lid sensor on top surface.

**Environmental conditions**


	Ambient temperature ( $t_{\text{ambient}}/^{\circ}\text{C}$ )	Uncertainty ( $U_{\text{ambient}}/^{\circ}\text{C}$ )	Limits ( $t_{\text{ambient}}/^{\circ}\text{C}$ )	Relative humidity (RH/%rh)	Uncertainty ( $U_{\text{rh}}/\text{%rh}$ )	Limits (RH/%rh)
Maximum	NA	1.0	27.0	NA	10.0	80.0
Minimum	NA	1.0	15.0	NA	10.0	20.0



### Temperature chart of thermal cycler (measured)



### Light intensity chart of optical system (emitted)



Protocol run time  
0 h : 28 m : 36 s



## Results

### Amplification curve

See Appendix A

### Cq results reported by real-time thermal cycler

Sensor	Cq	Specifications	Pass/Fail
1	17.50	NA	NA
2	17.65	NA	NA
3	17.15	NA	NA
4	17.68	NA	NA
5	17.36	NA	NA
6	17.19	NA	NA
7	17.43	NA	NA
8	17.68	NA	NA
9	17.40	NA	NA
10	17.15	NA	NA
11	17.58	NA	NA
12	17.44	NA	NA
13	17.37	NA	NA
14	17.39	NA	NA
15	17.21	NA	NA
<b>Average</b>	17.41		
<b>Standard deviation</b>	0.18		
<b>Uniformity (max - min)</b>	0.53		

Color filter: FAM

### Recommended control positions

Positive      A4    H1  
 Negative     D7    A10



## Appendix A

### **DUT: Device under test**

The device under calibration which is the real-time thermal cycler.

### **CU: Calibration Unit**

The measurement system which is used for execution of the calibration (thermal and/or optical).

### **Cq: Amplification curve**

Amplification curve = Quantification cycle. Cycle in which real-time PCR growth curve crossed the threshold.

During the amplification curve simulation, the real-time thermal cycler detects the light which is emitted by the optical calibration system and calculates Cq values for each channel. The light intensity is identical for each channel and gradually increased from 20 % to 100 %. The emitted light is independent of the temperature reached by each individual well.

The Cq value of each channel is determined by the DUT. In an idealized DUT the reported Cq values of all channels are identical and the uniformity is equal to 0 (zero).

A delta Cq of 1 represents a factor 2 difference in the DNA amount.

Generic reference value for each channel:  $-1 \leq Cq_{avg} \leq 1$   
(value can be used for normalization)



## Terminology

Uncertainty [°C]	Parameter associated with the result of a measurement that characterizes the dispersion of the values that could reasonably be attributed to the quantity intended to be measured.
Cq	Quantification cycle. Cycle in which real-time PCR growth curve crossed the threshold.
Uniformity [°C]	Maximum temperature minus minimum temperature at a specific time.
Uniformity Cq	Maximum Cq minus minimum Cq as determined by the DUT.



#### Calibration information

Calibration number	OA75012402-BY73
Software version	Cyclertest Measurement Software 1.0.5.23629
Engineer	Stephanie Bracht (Kalibrierlabor)
Data filename	A75012402-BY73_20260615142447.cl2
Sample interval	250ms
QMS document	TBD
Measurement system uncertainty	0.25 °C (k=2)

#### Equipment information (CU)

Probe serial	A75012402
Probe Type	GSOC Optical 96v15Plus4
Probe calibration date	31-Jul-2025
Probe calibration expire date	31-Jul-2026
Hardware box serial	A75012402
Hardware box version	0x0625
T-RH device serial	N/A
T-RH device calibration date	N/A

#### Disclaimer

MTAS® and Driftcon® are registered trademarks of CYCLERtest BV.  
All other mentioned (thermal cycler) instruments, block names and brand names are (registered) trademarks of their respective manufacturers or holders.  
The (optical) calibration methodology and hardware are covered by patents owned by CYCLERtest BV.

#### License of product usage

This product and its software are licensed to be used as calibration and/or verification tool. Paid and/or third-party calibration services with this product(s) is licensed on an annual base. License violation can lead to penalties.  
Please report license violation to support@cyclertest.com.  
© Layout Copyright protected 2024 and onward, CYCLERtest BV, The Netherlands.  
All rights reserved.



#### Control code

Go to <https://core.cyclertest.com/code/> to verify the data on this certificate.  
Code: KS5CZ-BZSX6-3E2RR-HDN8M-UYYHJ

Or scan the QR code to go to the page directly.